


World-beating  
sensitivity

世界最高  
レベル  
の感度



# CHEMILUMINESCENCE ANALYZER

Ultraweak luminescence detector systems



 東北電子産業株式会社

Tohoku Electronic Industrial Co., Ltd.



# Overview of Chemiluminescence Analyzers

## CLA-FS5

Using a photomultiplier tube (PMT), this is our most sensitive luminescence measurement device.




Typical samples




Solids, liquids and powders can also be measured, as long as they fit into a dish with a diameter of 50mm or 20mm.

Luminescence detection range




## CLA-IMG

A built-in ultrasensitive CCD camera enables this device to identify oxidised areas and perform measurements on multiple samples simultaneously.




## CLA-SP3



This is an instantaneous photometric weak-emission spectrum measurement device combining a high-sensitivity CCD camera with a bright spectrometer.




Chemiluminescence spectra



# Auto-oxidation mechanism and luminescence phenomena



## Data analysis method during heating measurement




The graph on the left shows typical CL behaviour during heating measurement. As the sample is heated, the peroxide decomposes, and CL from the excited carbonyl increases, resulting in a peak (the first peak). This corresponds to the amount of peroxide at that point. The oxidation reaction is then accelerated by heating in air or oxygen, and eventually the CL reaches a steady state. The intensity at this time is termed the steady-state luminescence intensity (*I<sub>s</sub>*). In the sample to which stabiliser has been added, the stabiliser is consumed, the steady state of the oxidation reaction is disrupted, and the amount of radicals in the sample increases, resulting in the appearance of significantly higher luminescence (the second peak). This point is called the oxidation induction time (*OIT*). The *OIT* can be used to evaluate the oxidative stability of the sample. Also, since *I<sub>s</sub>* is the steady state of radical extinction and formation within the sample, it represents the rate of radical generation, and this value can also be used to evaluate the oxidative stability of the sample.

## Measurement Modes

### CLA-FS5


CL Intensity measurement



The amount of luminescence at all wavelengths is measured for each exposure time setting (0.1 sec, 1 sec, 10 sec). The degree of oxidation is determined by parameters such as the height, slope, time and total area of the luminescence peak.


### CLA-IMG

CL Image measurement




### CLA-SP3


CL Spectrum measurement



CL Spectrum measurement



Spectrum measurement is performed while the spectral filter (high-pass filter) built into the device rotates automatically.







Instantaneous spectrum measurement can be performed with high resolution of 1nm or less.

Numerical values can be displayed within the selected range of the image.

# Measurement Examples (Polymers)

## Weather resistance evaluation

|                         |  |
|-------------------------|--|
| Samples                 | Polypropylene subjected to exposure testing and acceleration testing, with and without additives (HALS, UVA) |
| Exposure conditions     | 50MJ (approximately 2 months), 100MJ (approximately 4 months), JIS K 7219, exposure tests performed in Osaka |
| Acceleration conditions | 50MJ (equivalent to 177 hours), 100MJ (equivalent to 353 hours), JIS K 7350-4, Sunshine Weather Meter        |
| Measurement conditions  | 160°C, nitrogen, CLA-FS4   |






The "with additives" exposure-tested samples showed only a slight difference, but the acceleration-tested samples showed an increase in the amount of luminescence as they degraded, indicating a small difference in oxidative degradation at the very beginning.

In the tensile test, a difference in the value was first seen in the "without additives" acceleration-tested sample oxidised at 100MJ.

Samples were supplied by the Japan Chemical Innovation and Inspection Institute (JCII), along with acceleration test, exposure test and physical property test support.

## Evaluation of recycled materials

|                        |  |
|------------------------|--|
| Samples                | Polypropylene (PP) pellets manufactured at different extrusion temperatures and cycle counts |
| Extrusion conditions   | Temperatures: 230°C, 300°C; cycle counts: 0, 1, 3, 5   |
| Measurement conditions | 150°C, nitrogen, CLA-FS4   |




The higher the recycling cycle count, the higher the amount of luminescence shown; slight oxidation due to recycling could be detected. In the physical property tests (the flexure test and tensile test), hardly any difference was observed.

# Measurement Examples (Polymers)

## Evaluation of oxidation induction time (OIT)


|                        |   |
|------------------------|---|
| Samples                | Polypropylene (PP) pellets with different concentrations of additive (Irganox 1010) |
| Measurement conditions | CLA-FS4: 200°C, oxygen; CLA-IMG: 200°C, oxygen                                      |

OIT data measured using CLA-FS4




Luminescence image measurement

OIT image measured using CLA-IMG




Amount of luminescence

Large Small




Luminescence image measurement

|                        |  |
|------------------------|--|
| Samples                | Polyamide pellets<br>① Virgin: Unoxidised, without additive<br>② R1: Recycled, without additive<br>③ 1098_R1: Irgafos 1098, recycled<br>④ 412S_R1: AO-412S, recycled |
| Measurement conditions | CLA-IMG: 200°C, oxygen   |



## Luminescence spectrum during thermal oxidation

|                        |                            |
|------------------------|----------------------------|
| Samples                | Polypropylene (PP) pellets |
| Measurement conditions | 200°C, oxygen, CLA-SP3     |




With oxidation, a long wavelength shift was observed in the peak position in the 400nm range, and an increase in intensity was observed in the long wavelength region of 550nm and above.

## Simultaneous measurement of 100 samples


|                        |                            |
|------------------------|----------------------------|
| Samples                | Polypropylene (PP) pellets |
| Measurement conditions | 200°C, oxygen, CLA-100     |

●: PP; ○: PP + antioxidant



OIT measurement was carried out simultaneously on 100 samples. In the PP without the antioxidant, the OIT was shorter, and the OIT time course could be detected with good reproducibility, without being affected by adjacent samples.



CL time course for each sample



# Measurement Examples (Polymers)

## Evaluation of thermal oxidation of rubber

|                        |                                  |
|------------------------|----------------------------------|
| Samples                | Natural rubber with carbon black |
| Heat processing        | 100°C, 72 to 312 hours           |
| Measurement conditions | 160°C, oxygen, CLA-FS3           |




The longer the thermal oxidation time, the higher the amount of luminescence (see graph above); the integrated amount of luminescence at 1,800sec showed a high positive correlation with the thermal oxidation time (see graph on left).


## Comparison with infrared (IR) absorption measurement

|                        |                                       |
|------------------------|---------------------------------------|
| Samples                | Polypropylene powder                  |
| Degradation conditions | Heating at 160°C for 10 to 30 minutes |
| Measurement conditions | 160°C, nitrogen, CLA-FS4              |

### FT-IR



### CL




With the infrared absorption measurement method, a peak derived from the carbonyl group was visible in the sample after 30 minutes of heating (see graph on left), but with the CL method, an increase in luminescence was observed after 10 minutes of heating.

Samples supplied by: Sumitomo Chemical Co., Ltd.


## Magnified image measurement using CCD camera

|                        |   |
|------------------------|---|
| Samples                | Polycarbonate, thickness: 5mm   |
| Degradation conditions | Metal Weather exposure tester   |
| Measurement conditions | Using a SAICAS system, oblique cutting was performed to a depth of 25 $\mu\text{m}$ , and CL imaging measurement was performed on the cut surface. Laser irradiation (375nm, 10mW), CLA-IMG |



The more luminous (white) the sample, the more the oxidative degradation has progressed. Gradation was seen from the surface layer to the inner part.

### CL image from top of sample




Slides supplied by: Daipia Wintes Co., Ltd.  
Data supplied by: C. I. Takiron Corporation

# Measurement Examples (Polymers)


## Calculation of activation energy

|                        |                                      |
|------------------------|--------------------------------------|
| Samples                | Polypropylene (PP) pellets           |
| Measurement conditions | 100°C, 120°C, 160°C, oxygen, CLA-FS4 |



|         | 100°C  | 120°C  | 140°C  |
|---------|--------|--------|--------|
| 1/T     | 0.0026 | 0.0025 | 0.0024 |
| Is      | 191.62 | 239.12 | 331.73 |
| Ln (Is) | 3.00   | 4.21   | 5.08   |

- Oxidation of the samples was accelerated under oxygen flow at each temperature condition, and the average value was calculated for the value (Is) where luminescence stabilized after the first peak.
- Ea (activation energy) was determined from the slope of the equation of the approximation curve, with LN (Is) plotted as the vertical axis, and 1/T (absolute temperature) as the horizontal axis.



Slope (-8018.8) x gas constant = **66.7kJ/mol**





Ea can be calculated by measuring CL under multiple temperature conditions.

Samples supplied by: Sumitomo Chemical Co., Ltd.



## Lifespan prediction

|                        |  |
|------------------------|--|
| Samples                | Polypropylene (PP) pellets   |
| Measurement conditions | 50 to 250°C, oxygen<br>0.2°C/min, 0.4°C/min, 0.8°C/min,<br>CLA-FS4 |
| Analysis software      | AKTS - Thermokinetics  |

Reaction kinetics analysis program  
Friedman differential equivalent method  
System improving on the (Ozawa) integral equivalent method



- CL data
- Time-course graph of integrated value over time for amount of luminescence  
Reaction progress is represented as an integrated value over time. The integrated amount of luminescence at the point at which the temperature reached 160°C (melting point or lower) from the start of measurement at 0.8°C/min was taken to be the reaction end point. The reaction progress at this point was taken to be 1, and regarded as the lifespan.
- The Ln (dα/dt) for α (reaction progress) was determined from a graph of the relationship between the reciprocal of each temperature and the logarithm of the reaction speed. This slope is the Ea (activation energy); the Ea for each α can be determined.
- Activation energy
- Lifespan prediction results at constant temperatures (30°C, 50°C, 70°C, 90°C, 110°C) (vertical axis: reaction progress; horizontal axis: time)
- Lifespan prediction (bar graph)




This suggests that lifespan can be predicted by means of CL measurement under multiple temperature-elevation measurements, without acceleration testing.

Samples supplied by: Sumitomo Chemical Co., Ltd.;  
Analysis support: Palmetrics Corporation

# Measurement Examples (Food)

## Measurement of rapeseed oil (1)


|                        |   |
|------------------------|---|
| Samples                | Rapeseed oil<br>New, refrigerated for 2 years |
| Measurement conditions | 50 to 250°C, nitrogen, CLA-FS4                |



The amount of luminescence was higher in the sample refrigerated for 2 years than in the new sample, and multiple luminescent components were observed by means of temperature-elevation measurement.

## Measurement of rapeseed oil (2)

|                        |                         |
|------------------------|-------------------------|
| Samples                | 5g rapeseed oil         |
| Measurement conditions | 150°C, nitrogen, CLA-ID |




The CL integrated value and the POV value showed a high correlation.

R. Ushuki, Nippon Shokuhin Kogyo Gakkaishi 32 (1), 74 (1985)

## Luminescence of beer

|                        |                                  |
|------------------------|----------------------------------|
| Samples                | 1.2ml beer                       |
| Degradation conditions | Stored at 30°C for up to 6 weeks |
| Measurement conditions | 60°C, CLA-ID                     |




The longer the storage period, the more the CL increased.

H. Kaneda et al., Journal of Food Science, 55 (5), 1361-1364, 1990

## Measurement of catechins in human blood

|                        |   |
|------------------------|---|
| Samples                | Catechin extract in plasma  |
| Mobile phase           | Methanol-water (2:8, v/v, containing 0.1% phosphoric acid), 1.0ml/min   |
| Reagents               | ① 8.0M acetaldehyde in 50mM phosphate buffer at pH 7.4, containing HRP 108mg/L, 3.0ml/min<br>② 8.8M H <sub>2</sub> O <sub>2</sub> , 1.0ml/min |
| Measurement conditions | CLA-FL, HPLC system   |

Detection of EGCg luminescence in human plasma, before and after ingestion of EGCg




Nakagawa, K. and Miyazawa, T.: Analytical Biochemistry, 248, 41-49, 1997

Luminescence peaks of epigallocatechin gallate (EGCg) were detected in plasma 60 minutes after ingesting an EGCg capsule.

## Measurement of ATP

|                        |                                       |
|------------------------|---------------------------------------|
| Samples                | ATP reagent manufactured by Cosmo Bio |
| Measurement conditions | Room temperature, air, CLA-IDsp       |

ATP measurement by means of L-L reaction (using Cosmo Bio's Kinshiro)




Luminescence up to about  $1 \times 10^{-13}$  M showed good linearity, with a correlation coefficient of 0.97.

## Luminescence of cookies

|                        |                                    |
|------------------------|------------------------------------|
| Samples                | Cookies (deep-fried confectionery) |
| Degradation conditions | 254nm, irradiation for 0 to 1 hour |
| Measurement conditions | 100°C, nitrogen, CLA-ID            |

Unirradiated      Irradiated for 1 hour




The amount of luminescence increased after 1 hour of light irradiation.

Samples supplied by: Kochi Prefecture Paper Technology Center

## Measurement of coffee

|                        |                                 |
|------------------------|---------------------------------|
| Samples                | Columbian medium-roasted beans  |
| Measurement conditions | 80°C nitrogen → oxygen, CLA-FS4 |






Immediately after grinding, the amount of luminescence was higher and the fragrance and flavour were also better. The amount of luminescence decreased with storage time.

Samples supplied by: La Coet

# Measurement Examples (Biochemical)

## Measurement of phospholipid hydroperoxides (PCOOH) in blood

|                        |   |
|------------------------|---|
| Samples                | Catechin extract in plasma  |
| Mobile phase           | 2-propanol-methanol-water (135:45:20, v/v/v)  |
| Reagents               | 10mg of cytochrome c and 2mg of luminol dissolved in 1L of 50mM borate buffer solution                              |
| Sample                 | Photo-oxide of L- $\alpha$ -phosphatidylcholine, $\beta$ -oleoyl- $\gamma$ -palmitoyl (C18:1, [cis]-9/C16:0, SIGMA) |
| Measurement conditions | CLA-FL HPLC system (column: SIL-NH2)  |





CL-HPLC chromatograms were obtained for normal subjects a and b. The amount of PCOOH for subject a (several hundred femtomoles) was less than for subject b (several picomoles), and could be detected with good reproducibility. Lipid peroxides in human blood are an indicator of oxidative stress in the body.

Guidance provided by: Professor Teruo Miyazawa, Tohoku University Graduate School of Agriculture

## Ultra-sensitive fluorescence measurement

|                        |   |
|------------------------|---|
| Samples                | Haematoporphyrin                                      |
| Measurement conditions | LD 405nm + HP 600nm<br>Room temperature, air, CLA-FS4 |






In contrast to the method using a general fluorescence spectrophotometer, this method enabled a calibration curve to be obtained to about 20fmol.


## Measurement of tablets

|                        |  |
|------------------------|--|
| Samples                | Tablets of the same kind   |
| Degradation conditions | Experiment 1: All new tablets<br>Experiment 2: photodegraded tablets (1 week under indoor diffused light, 2 weeks under 4000Lux)                           |
| Measurement conditions | Experiment 1: 150°C, oxygen, exposure for 1 min, sensitivity: 255, CLA-IMG<br>Experiment 2: 150°C, nitrogen, exposure for 1 min, sensitivity: 255, CLA-IMG |

With and without a formulation of different constituents (Formulation A) placed in the middle



The amount of luminescence was higher, and oxidation was more prone to occur, when Formulation A was present.




Although the tablets had the same main constituents, a difference was seen in the rate of increase.

# Specifications

| Product name                            | CLA-FS5   | CLA-ID5  |
|---|---|--|
| Photograph                              |    |  |
| Detection method                        | Single photon counting method<br>using a photomultiplier tube   |  |
| Detection wavelength                    | 300nm to 650nm (centre wavelength: 420nm)   |  |
| Cooling method                          | Primary cooling: Peltier element; Secondary cooling: water cooling  |  |
| Measurement items                       | ① Luminescence intensity (counts per second)<br>② Luminescence spectrum<br>(380nm to 660nm/20nm resolution)   | Luminescence intensity (counts per second)   |
| Minimum measurement time<br>(Gate time) | 0.1 sec, 1 sec, 10 sec  |  |
| Spectral filters                        | 15, built-in<br>(380nm to 660nm: every 20nm)  | None   |
| Touch panel<br>display items            | ① Amount of luminescence, ② Sample chamber temperature, ③ Sample chamber temperature setting, ④ Status, ⑤ Gate time, ⑥ Alarm, ⑦ Detail, ⑧ Sample chamber status (open/closed), ⑨ Shutter status (open/closed) |  |
| Communication functionality             | 1 USB port (used by dedicated software)   |  |
| Dimensions, weight                      | 523.5mm (W) x 411.5mm (D) x 547mm (H)<br>Approx. 60kg   | 310mm (W) x 420mm (D) x 524mm (H)<br>Approx. 35kg                                  |

# Sample Chamber Specifications

| Product, model         | Sample Chamber<br>(Heating Type to 220°C)<br>CLS-ST5   | Sample Chamber<br>(Non-isothermal<br>Type)<br>CLS-SH2  | Sample Chamber<br>(Mixing Type)<br>CLS-MX5  | Sample Chamber<br>(Flow Type)<br>CLS-FL2   |
|------------------------|--|--|---|--|
| Maximum sample size    | 50mm diameter x 10mm (H)   | 20mm diameter x 5mm<br>(H)   | 50mm diameter x 10mm<br>(H)   | Flow tube<br>bore: 0.5mm   |
| Heating temperature    | Room temperature to 220°C  | Room temperature to<br>350°C   | Room temperature to<br>100°C  | Room temperature to<br>50°C  |
| Functionality included | Atmosphere replacement   | Non-isothermal<br>functionality<br>Atmosphere replacement  | Atmosphere replacement<br>Sample agitation<br>Reagent injection   | 2 injection ports<br>1 drainage port   |
| Dimensions, weight     | <br>221mm (W) x 357mm (D) x<br>121mm (H)<br>Approx. 4kg | <br>221mm (W) x 357mm (D) x<br>121mm (H)<br>Approx. 4kg | <br>221mm (W) x 357mm (D) x<br>121mm (H)<br>Approx. 4kg | <br>221mm (W) x 357mm (D) x<br>121mm (H)<br>Approx. 2kg |

# Specifications

| Product name                | CLA-IMG4   | CLA-SP3  |
|-----------------------------|--|--|
| Photograph                  |     |  |
| Detection method            | Back-illuminated frame-transfer CCD camera   |  |
| Detection wavelength        | 400 to 800nm (centre wavelength: 600nm)  |  |
| Cooling method              | Air cooling  |  |
| Number of effective pixels  | 1024 x 1024  | 1600 x 200   |
| Resolution                  | Vacuum resolution: approx. 150 $\mu$ m × 150 $\mu$ m<br>(Option: approx. 10 $\mu$ m) | Wavelength resolution: 1nm   |
| Measurement items           | Luminescence image<br>Luminescence intensity (within image selection range)          | Luminescence spectrum measurement  |
| Exposure time               | 30ms to 120min   | 0.01 to 10,000sec  |
| Lens                        | 25mm, F0.95 (C mount)  | Incidence slit width: 0.1/0.5/1.0mm  |
| Built-in shutter            | Built-in mechanical shutter  | None   |
| Communication functionality | IEEE1394b  | USB  |
| Dimensions, weight          | 310mm (W) x 446mm (D) x 775mm (H)<br>Approx. 30kg                                    | 310mm (W) x 420mm (D) x 524mm (H)<br>Approx. 35kg                                  |

## Sample Chamber Specifications

| Product, model                | Sample Chamber<br>(Laser-induced<br>Fluorescence Type)<br>CLS-LA1   |
|-------------------------------|---|
| Maximum sample size           | 50mm diameter x 10mm (H)  |
| Heating temperature           | Room temperature to 100°C   |
| Laser light-source wavelength | 375nm or 405nm  |
| Laser output and stability    | 0.1 to 20mW<br>At 5 to 20mW: $\pm$ 1%<br>At 0.1 to 5mW: $\pm$ 5%  |
| Dimensions, weight            | <br>221mm (W) × 357mm (D) × 121mm (H)<br>Approx. 4kg |

## Accreditations and awards

- 2006: Certified by the Ministry of Economy, Trade and Industry (METI) of the Government of Japan as one of Japan's 300 Most Vibrant Monozukuri (Manufacturing) Small and Medium Enterprises
- 2009: Received Miyagi Sugure Mono ("Miyagi outstanding product") accreditation under a promotional project spearheaded by Miyagi Prefecture and other bodies
- 2012: Received the Tohoku Bureau of Economy, Trade, and Industry Director-General's Award, one of the Monozukuri Nippon Grand Awards
- 2014: Received the First Technology Advancement Award conferred by the Japan Society of Polymer Processing
- 2017: Certified by METI as a Company Driving Regional Growth
- 2019: Certified by the Kawasaki Monozukuri (Manufacturing) Brand Promotion Council as a Kawasaki Monodukuri Brand



2018: Chemiluminescence methods named in a newly-published Japanese Industrial Standard (JIS)

### K 7351

"Sensitive Measurement Method of Peroxide In Plastics By Detecting Ultra-Weak Photon Emission"





<http://www.tei-c.com>

## **Tohoku Electronic Industrial Co., Ltd.**

Head Office: 2-14-1 Mukaiyama, Taihaku-ku, Sendai, Miyagi 982-0481, Japan

Rifu Office: 6-6-6 Shirakashi-dai, Rifu-cho, Miyagi 981-0134, Japan

Tokyo Office: 203, Lapole-Shinmaruko, 2-897 Shinmaruko Higashi, Nakahara-ku, Kawasaki 211-0004, Japan

Kyoto Lab.: 4F-B, Kohei Bldg., 717 Uematsu-cho, Kawaramachi-dori Matsubara-sagaru, Shimogyo-ku, Kyoto, Kyoto 600-8028, Japan

Note : "Chemiluminescence Analyzer" is a trademark of Tohoku Electronic Industrial Co., Ltd

Note : Please be aware that the information given in this document is subject to change without notice, for the purposes of improvement et cetera.

Note : Unauthorized reproduction of the content of this document in whole or in part is forbidden.