



財團法人全國認證基金會  
Taiwan Accreditation Foundation

# 認證證書

(證書編號：L3481-230717)

茲證明

大泉機器廠有限公司

泵浦測試實驗室

新北市土城區中山路 83 號

為本會認證之實驗室

認證依據：ISO/IEC 17025：2017；CNS 17025：2018

認證編號：3481

初次認證日期：一百零七年四月十九日

認證有效期間：一百一十年五月十七日至一百一十三年五月十六日  
止

認證範圍：測試領域，如續頁

董事長

連錦漳



掃描確認真偽

中華民國一十二年七月十七日

認 證 編 號：3481

實驗室主管：張建業

■ 16.05 機械儀器設備

水泵浦

M996 揚程、流率、動力及效率

ISO 9906

CNS 16017

ISO 1438

揚程範圍：

壓力計範圍：(-1 to 25) bar

流率範圍：

四角堰：(288 to 13072) L/min

三角堰：(100 to 1481) L/min

電磁流量計法：(30 to 4242) L/min

泵驅動機輸入功率：(0.075 to 100) kW

性能試驗允收等級 2 級及 3 級

報告簽署人：張怡泰；張建業；張義杰

(以下空白)





財團法人全國認證基金會  
Taiwan Accreditation Foundation

## Certificate of Accreditation

(Certificate No : L3481-230717)

This is to certify that

**TA CHUANG MACHINE FACTORY CO., LTD.**

**Pump Test Laboratory**

NO.83, Zhongshan Rd., Tucheng Dist., New Taipei City 236, Taiwan (R.O.C.)

**is accredited in respect of laboratory**

**Accreditation Criteria** : ISO/IEC 17025:2017 ; CNS 17025:2018

**Accreditation Number** : 3481

**Originally Accredited** : April 19, 2018

**Effective Period** : May 17, 2021 to May 16, 2024

**Accredited Scope** : Testing Field, see described in the Appendix



Scan to verify

*Ching-Chang Lien*

Ching-Chang Lien  
President, Taiwan Accreditation Foundation  
July 17, 2023

Accreditation Number : 3481

Laboratory Head : CHANG, Chien-Yeh

■ 16. 05 Machine and Equipment

Water Pumps

M996 Head, Flow Rate, Power and Efficiency

ISO 9906

CNS 16017

ISO 1438

Head Range:

Pressure Gauge Range: (-1 to 25) bar

Flow Rate Range:

Rectangular Weir: (288 to 13072) L/min

Triangular Weir: (100 to 1481) L/min

Electromagnetic Flow Meter Method: (30 to 4242) L/min

Power transmitted to the pump by its driver: (0.075 to 100) kW

Performance Acceptance Test Grades 2 & 3

Approval Signatory: CHANG, Yi-Tay; CHANG, Chien-Yeh; CHANG, I-Chieh

(Null below)

